

# CALsys -35/200 Evaluation Report

An evaluation report of the CALsys -35/200 Liquid Calibration bath

Manufactured by Tempsens Instruments (II) Pvt Ltd

## INTRODUCTION

The is the latest version of Tempsens makes most popular High Accuracy liquid calibration Bath. It works over the temperature range -35 Deg C to 200 Deg C.

At Tempsens it is our earnest desire to present for our customer's consideration as much useful information as possible and to this end we have spent a substantial amount of time evaluating our products.

The results of the evaluation of Accuracy liquid calibration Bath can be presented in many formats some of which will give an optimistic or indeed a pessimistic view of how the products operate. The performance of the bath will vary depending on liquid type, stirring speed and other outside influences.



**CALsys -35/200**

## A. Radial temperature homogeneity:

### What is Radial Temperature Homogeneity and why it is important to measure

Radial uniformity refers to temperature differences between wells of the bath. This non uniformity is strongly influenced by the difference between the bath and ambient temperature. A larger temperature difference from ambient will result in a larger potential temperature calibration error. Therefore, radial in homogeneity should be measured at extremes (relative to ambient temperature) in an instrument's temperature range.

### Test method:


The temperature differences between the zones in the individual bores provided for the measurements are measured with one or several suitable thermometers at three different temperature representative of the field of application and covering the extreme temperature which may occur. If there is only one bore no measurement has to be carried out.

For CALsys -35/200, we consider -35 Deg C, 50 Deg C and 200 Deg C respectively. As example at -35 Deg C two RTD (designed for small steam conduction) were placed in each of the holes. Measurements were recorded and then the probes were interchanged between the two pockets and repeat measurements made. The temperature Difference was calculated to remove the small offsets between the two probes.

For calibrators having fewer than four wells, it may be necessary to determine differences by cyclic exchange. The difference between two wells with two thermometers may be determined with the following Formula:

$$\text{Temperature Difference} = [((P1W1 - P1W2) + (P2W1 - P2W2)) / 2]$$

Note: P1 = probe 1, W1 = well 1 and so on. P1W1 is read as the value of probe 1 in well 1.



**Radial temperature Homogeneity at CALsys -35/200**

Temp (Deg C)	Sensor	Hole 1	Hole 2
-35°C	RTD Sr.No 1354	-35.343	-35.356
	RTD Sr.No 1338	-35.563	-35.507
Radial Uniformity : $\pm 0.0215$			

Temp (Deg C)	Sensor	Hole 1	Hole 2
50°C	RTD Sr.No 1354	49.641	49.723
	RTD Sr.No 1338	49.601	49.432
Radial Uniformity : $\pm 0.0445$			

Temp (Deg C)	Sensor	Hole 1	Hole 2
200°C	RTD Sr.No 1354	199.462	199.568
	RTD Sr.No 1338	199.462	199.485
Radial Uniformity: $\pm 0.0645$			

**B. Temperature Stability**


Temperature stability is measured with a thermometer and readout with adequate sensitivity and resolution to measure the control fluctuations in the bath. A typical time period for stability measurements of a bath is about 30 minutes at any specific temperature. Other time periods may be applied depending on how the calibrator is to be used. Temperature stability may vary at different temperatures. The instrument should be characterized over its range, and typically three sets of stability measurements are adequate. Baths that are heated only (that is they utilize no cooling systems to achieve below-ambient temperatures) are measured at their maximum and minimum temperatures and at the midpoint of their ranges. Stability measurements for bath are made at their maximum and minimum temperatures as well as near room temperature. Specific temperatures of interest by the user may also be incorporated.

**Test Method:**

Stability is the measure of the temperature deviations over the measurement period, after temperature control has stabilized. The stability data can be viewed in two ways (Stability graph). What may be called “peak” stability is often evaluated as plus or minus (±) one-half the difference between of the maximum and minimum values of the data set:

**Peak Stability = ± (Tmax – Tmin) / 2.**

Temp Set Point	Maximum	Minimum	Peak Stability
-35 Deg C	-35.544	-35.564	0.01
50 Deg C	50.247	50.207	0.02
200 Deg C	200.842	200.766	0.038



**C.HEAT UP TIME and COOL DOWN TIME**

**HEAT UP TIME**

Amb.to 200°C 60 min ; -35°C to Amb. 30 min

**COOL DOWN TIME**

Amb. to -35° C 60 min ; 200°C to 100°C 120 min

Heat Up Time	
Time (Min)	Temp (Deg C)
0	22.920
1	23.451
3	25.631
5	46.785
7	77.965
10	107.167
20	157.906
40	189.266
60	199.962

Cool Down Time	
Time (Min)	Temp (Deg C)
0	23.336
1	22.703
3	17.148
5	1.039
10	-8.703
20	-19.406
40	-24.727
50	-33.336
60	-35.078

